

Peroxide CHEMets®

0 - 1 & 1 - 10 ppm

Safety Information

Read MSDS before performing this test procedure. Wear safety glasses.

Test Procedure

1. Fill the sample cup to the 25 mL mark with the sample (fig 1).
2. Place the CHEMet ampoule in the sample cup. Snap the tip by pressing the ampoule against the side of the cup. The ampoule will fill leaving a small bubble to facilitate mixing (fig 2).
3. Mix the contents of the ampoule by inverting it several times, allowing the bubble to travel from end to end each time. Wipe all liquid from the exterior of the ampoule. Test results should be obtained **within one minute** after snapping the ampoule tip.
4. Use the appropriate comparator to determine the level of hydrogen peroxide in the sample. If the color of the CHEMet ampoule is between two color standards, a concentration estimate can be made.




Figure 1




Figure 2




Figure 3

- a. Place the CHEMet ampoule, flat end downward into the center tube of the low range comparator. Direct the top of the comparator up toward a source of bright light while viewing from the bottom. Rotate the comparator until the color standard below the CHEMet ampoule shows the closest match (fig 3).

- b. Hold the high range comparator in a nearly horizontal position while standing directly beneath a bright source of light. Place the CHEMet ampoule between the color standards moving it from left to right along the comparator until the best color match is found (fig 4).




Figure 4

Test Method


The Peroxide CHEMets®¹ test method employs the ferric thiocyanate chemistry.² In an acidic solution, hydrogen peroxide oxidizes ferrous iron. The resulting ferric iron reacts with ammonium thiocyanate to form ferric thiocyanate, a red-orange colored complex, in direct proportion to the hydrogen peroxide concentration. Results are expressed in ppm (mg/Liter) H₂O₂. Various oxidizing agents such as ozone, ferric ions and cupric ions will produce high test results.

1. CHEMets is a registered trademark of CHEMetrics, Inc. U.S. Patent No. 3,634,038
2. D. F. Boltz and J. A. Howell, eds., Colorimetric Determination of Nonmetals, 2nd ed., Vol. 8, p. 304 (1978)

Reorder Information

Cat. No.

<i>Test Kit, complete</i>	<i>K-5510</i>
<i>Refill, 30 CHEMet ampoules</i>	<i>R-5510</i>
<i>Sample Cup, 25 mL, package of six</i>	<i>A-0013</i>
<i>Comparator, 0-1 ppm</i>	<i>C-5501</i>
<i>Comparator, 1-10 ppm</i>	<i>C-5510</i>



CHEMetrics, Inc., 4295 Catlett Road, Calverton, VA 20138-0214 U.S.A.
 Phone: (800) 356-3072; Fax: (540) 788-4856; E-Mail: orders@chemetrics.com
www.chemetrics.com Apr. 06, Rev. 3